


## Milton Regional Sewer Authority Wastewater Laboratory

<u>Test Parameter</u>	<u>Amount Necessary</u>	<u>Preservation Required</u>	<u>Cost/test</u>	<u>Accredited Method</u>
Ammonia as N*	500 mL	H <sub>2</sub> SO <sub>4</sub> to pH<2; 4°C	\$15.00	HACH 8038
BOD <sub>5</sub>	500 mL	4°C	\$23.00	SM 5210B
CBOD <sub>5</sub>	500 mL	4°C	\$23.00	SM 5210B
Fecal Coliform*	100 mL	4°C	\$20.00	SM 9222D
Nitrite as N	100 mL	4°C	\$15.00	HACH 8507
Total Phosphorus	100 mL	H <sub>2</sub> SO <sub>4</sub> to pH<2; 4°C	\$15.00	HACH 8190
Total Solids	100 mL/100 g	4°C	\$10.00	SM 2540G
Total Suspended Solids (TSS)	500 mL	4°C	\$10.00	SM 2540D

\*Samples are to be properly dechlorinated prior to preservation and delivery, if necessary.



Additional, non-accredited testing is also available by request

