


## Milton Regional Sewer Authority Wastewater Laboratory

| <u>Test Parameter</u>        | <u>Amount Necessary</u> | <u>Preservation Required</u>                | <u>Cost/test</u> | <u>Accredited Method</u> |
|------------------------------|-------------------------|---|------------------|--------------------------|
| Ammonia as N*                | 500 mL                  | H <sub>2</sub> SO <sub>4</sub> to pH<2; 4°C | \$15.00          | EPA 350.1                |
| BOD <sub>5</sub>             | 500 mL                  | 4°C   | \$23.00          | SM 5210B                 |
| CBOD <sub>5</sub>            | 500 mL                  | 4°C   | \$23.00          | SM 5210B                 |
| Fecal Coliform*              | 100 mL                  | 4°C   | \$20.00          | SM 9222D                 |
| Nitrite as N                 | 100 mL                  | 4°C   | \$15.00          | HACH 8507                |
| Total Phosphorus             | 100 mL                  | H <sub>2</sub> SO <sub>4</sub> to pH<2; 4°C | \$15.00          | SM4500-P E               |
| Total Solids                 | 100 mL/100 g            | 4°C   | \$12.00          | SM 2540G                 |
| Total Suspended Solids (TSS) | 500 mL                  | 4°C   | \$12.00          | SM 2540D                 |

\*Samples are to be properly dechlorinated prior to preservation and delivery, if necessary.



Additional, non-accredited testing is also available by request

Contact Genie Bausinger at 570-742-3424 or  
gbausinger@miltonregional.org for scheduling

